

REF 600-2	30x1.7mL	EXP 2025-12-17	LOT 393DD5F	更新日期： 2022-02
				1 / 1

批号：393DD5F				
项目\仪器\试剂\方法	单位	均值	+ / - 2 SD	
(Ca)钙				
Roche cobas b 121/123/221 Dedicated \ ISE (indirect)	mmol/L	1.12	0.9-1.34	
Siemens RAPIDPoint 400/500 Dedicated \ ISE (direct)	mmol/L	1.14	0.9-1.38	
Radiometer ABL Series Dedicated \ ISE (indirect)	mmol/L	1.09	0.87-1.31	
(Cl)氯				
IL Gem Premier 3000 series Dedicated \ ISE (direct)	mmol/L	94.3	76.1-112	
Siemens RAPIDPoint 400/500 Dedicated \ ISE (direct)	mmol/L	99.6	81.4-118	
Radiometer ABL Series Dedicated \ ISE (direct)	mmol/L	92.3	74.1-110	
(K)钾				
IL Gem Premier 3000 series Dedicated \ ISE (direct)	mmol/L	4.03	3.25-4.81	
Siemens RAPIDPoint 400/500 Dedicated \ ISE (direct)	mmol/L	4.11	3.33-4.89	
Radiometer ABL Series Dedicated \ ISE (direct)	mmol/L	4.09	3.33-4.85	
(Na)钠				
IL Gem Premier 3000 series Dedicated \ ICP-MS	mmol/L	145	121-169	
Siemens RAPIDPoint 400/500 Siemens \ ISE (indirect)	mmol/L	142	120-164	
Radiometer ABL Series Dedicated \ ISE (direct)	mmol/L	141	117-165	
Radiometer AQT Series Dedicated \ ISE (direct)	mmol/L	139	115-163	
(PCO2)二氧化碳分压				
IL Gem Premier 3000 series Dedicated \ Electrode	mmHg	35.9	30-41.8	
Roche cobas b 121/123/221 Dedicated \ Electrode	mmHg	39.6	32.2-47	
Siemens RAPIDPoint 400/500 Siemens \ Glucose oxidase (oxygen electrode)	mmHg	31.8	25.3-38.3	
Nova Stat Profile pHox series Dedicated \ Electrode	mmHg	32.6	26.7-38.5	
Radiometer ABL Series Dedicated \ Electrode	mmHg	40.6	32.8-48.4	
Radiometer AQT Series Dedicated \ Electrode	mmHg	42.6	34.8-50.4	
(pH)酸碱度				
IL Gem Premier 3000 series Dedicated \ Electrode	No Units	7.46	6.08-8.84	
Roche cobas b 121/123/221 Dedicated \ Electrode	No Units	7.43	6.01-8.85	
Siemens RAPIDPoint 400/500 Siemens \ Electrode	No Units	7.42	6.14-8.7	
Radiometer ABL Series Dedicated \ Electrode	No Units	7.49	6.13-8.85	
Radiometer AQT Series Dedicated \ Electrode	No Units	7.38	6-8.76	
(PO2)氧分压				
IL Gem Premier 3000 series Dedicated \ Electrode	mmHg	96.2	79-113	
Roche cobas b 121/123/221 Dedicated \ Electrode	mmHg	108	90-126	
Siemens RAPIDPoint 400/500 Dedicated \ Electrode	mmHg	122	101-143	
Nova Stat Profile pHox series Dedicated \ Electrode	mmHg	123	103-143	
Radiometer ABL Series Dedicated \ Electrode	mmHg	119	97-141	
Radiometer AQT Series Dedicated \ Electrode	mmHg	94.6	76.8-112	

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。